

## **Specification Sheet**

Specification Number: 8063H-300S Revision: 09

Date Issued: 23-Jun-23

**1. Part Number**: 8063H-300S

### 2. Photograph:



#### 3. Product Information:

Description: SteriWare® Powder Lance - 400mm – Sterile

Use: Single use, disposable powder sampler

Material of Construction: HDPE (High Density Polyethylene) - virgin

• Conforms to FDA CFR 177.1520

Conforms to EU Regulations 10/2011

• Conforms to EC Regulations 1935/2004

• BSE/TSE free

Method of Construction: Ultrasonic welding

Nominal Length:

Blade Length:

Diameter of Blade:

Nominal Weight of Sampler:

401mm

284mm

22mm

64g

Moulding Environment: Class 100,000 Cleanroom Welding & Primary Packing Environment: Class 10,000 Cleanroom

Individually Bagged? Yes (heat sealed PE/PET bag)
Method of Sterilisation: Gamma Irradiation (25-45kGy)

Quantity per Box: 20

Recommended Storage Conditions: Dry and ambient temperature

Shelf Life: 5 years from the date of manufacture
ATEX: This device has not been ATEX tested

#### **4. Change Control:** Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the SteriWare range please email your contact details to

change.control@sampling.com



## Sampling Systems Ltd.

4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com



# **Specification Sheet**

Specification Number: 8063H-300S Revision: 09 Date Issued: 23-Jun-23

**5. Country of Origin:** Made and packed in the UK.

**6. Repeated Use:** This is a single-use, disposable sampler. Repeated use will lead to

failure.

7. Approval: Signed: Name: Kuldeep Flora

Position: Quality Manager

Date: 23-Jun-23